

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052622\
 Data File : VN072642.D
 Acq On : 26 May 2022 21:06
 Operator : JC\MD
 Sample : N3029-02DL 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DSN-002-SEWER-OUTLET-002DL

Quant Time: May 27 05:35:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N052622W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 27 05:07:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.657	128	29851	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.963	114	156828	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.739	117	132323	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.434	65	83600	29.268	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	97.567%
60) 4-Bromofluorobenzene	12.727	95	58271	25.075	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	83.600%
63) Toluene-d8	10.439	98	233119	30.984	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.267%
Target Compounds						
					Qvalue	
15) Acetone	4.287	58	19304	49.346	ug/l	86
25) Chloroform	7.810	83	8021	1.811	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052622\
Data File : VN072642.D
Acq On : 26 May 2022 21:06
Operator : JC\MD
Sample : N3029-02DL 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DSN-002-SEWER-OUTLET-002DL

Quant Time: May 27 05:35:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N052622W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 27 05:07:24 2022
Response via : Initial Calibration

