

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
 Data File : VN077924.D
 Acq On : 26 May 2023 13:53
 Operator : JC\MD
 Sample : 02924-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-002-003-1

Quant Time: May 29 00:28:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	383534	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	703511	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	598481	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	188263	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	264601	46.109	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.220%		
35) Dibromofluoromethane	8.171	113	223329	48.682	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.360%		
50) Toluene-d8	10.571	98	849242	48.095	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.180%		
62) 4-Bromofluorobenzene	12.853	95	278572	44.669	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.340%		
Target Compounds					Qvalue	
16) Acetone	4.447	43	18499	10.246	ug/l	98
20) Methylene Chloride	5.277	84	13364	0.904	ug/l #	85
30) Chloroform	7.965	83	17910	1.913	ug/l	88
47) Bromodichloromethane	9.888	83	7500	1.084	ug/l #	98
52) Toluene	10.635	92	2610	0.212	ug/l	98
60) Dibromochloromethane	11.359	129	3737	0.740	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
Data File : VN077924.D
Acq On : 26 May 2023 13:53
Operator : JC\MD
Sample : 02924-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
001-002-003-1

Quant Time: May 29 00:28:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

