

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
 Data File : VN077926.D
 Acq On : 26 May 2023 14:41
 Operator : JC\MD
 Sample : 02924-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-002-003-3

Quant Time: May 29 00:29:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	397747	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	716655	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	627331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	200469	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	271081	45.550	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.100%
35) Dibromofluoromethane	8.177	113	228379	48.870	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.740%
50) Toluene-d8	10.576	98	901028	50.092	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.180%
62) 4-Bromofluorobenzene	12.859	95	285679	44.969	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.940%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	13592	7.259	ug/l	91
20) Methylene Chloride	5.289	84	13181	0.759	ug/l	96
30) Chloroform	7.977	83	19230	1.981	ug/l	99
47) Bromodichloromethane	9.894	83	10566	1.499	ug/l #	93
60) Dibromochloromethane	11.371	129	4725	0.918	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
Data File : VN077926.D
Acq On : 26 May 2023 14:41
Operator : JC\MD
Sample : 02924-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
001-002-003-3

Quant Time: May 29 00:29:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

