

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
 Data File : VN077929.D
 Acq On : 26 May 2023 15:53
 Operator : JC\MD
 Sample : 02945-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TWP03

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/30/2023
 Supervised By : Mahesh Dadoda 05/30/2023

Quant Time: May 29 00:30:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	404215	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	729554	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	627342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	186659	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	275680	45.581	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.160%		
35) Dibromofluoromethane	8.177	113	237128	49.845	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.680%		
50) Toluene-d8	10.577	98	896390	48.953	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.900%		
62) 4-Bromofluorobenzene	12.859	95	280877	43.431	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	86.860%		
Target Compounds						
19) Methyl tert-butyl Ether	5.812	73	6154m	0.400	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
Data File : VN077929.D
Acq On : 26 May 2023 15:53
Operator : JC\MD
Sample : 02945-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TWP03

Quant Time: May 29 00:30:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/30/2023
Supervised By :Mahesh Dadoda 05/30/2023

