

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
 Data File : VN077931.D
 Acq On : 26 May 2023 16:41
 Operator : JC\MD
 Sample : 02955-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TWP05

Quant Time: May 29 00:30:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	371228	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	675540	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	602489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	198646	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	261057	46.999	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.000%
35) Dibromofluoromethane	8.177	113	222128	50.425	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.840%
50) Toluene-d8	10.576	98	841772	49.646	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.300%
62) 4-Bromofluorobenzene	12.859	95	278350	46.482	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.960%
Target Compounds						
					Qvalue	
16) Acetone	4.459	43	3575	2.046	ug/l #	87
19) Methyl tert-butyl Ether	5.812	73	6665	0.472	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
Data File : VN077931.D
Acq On : 26 May 2023 16:41
Operator : JC\MD
Sample : 02955-13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TWP05

Quant Time: May 29 00:30:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

