

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
 Data File : VN077932.D
 Acq On : 26 May 2023 17:05
 Operator : JC\MD
 Sample : 02955-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP07

Quant Time: May 29 00:31:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	307585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	556760	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	481115	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	144439	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	212601	46.195	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.380%
35) Dibromofluoromethane	8.177	113	182064	50.148	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.300%
50) Toluene-d8	10.577	98	680023	48.662	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.320%
62) 4-Bromofluorobenzene	12.859	95	215757	43.716	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.440%
Target Compounds						
19) Methyl tert-butyl Ether	5.812	73	5335	0.456	ug/l	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
Data File : VN077932.D
Acq On : 26 May 2023 17:05
Operator : JC\MD
Sample : 02955-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP07

Quant Time: May 29 00:31:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

