

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
 Data File : VN077935.D
 Acq On : 26 May 2023 18:17
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: May 29 00:32:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	361271	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	652728	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	560900	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	176120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	250257	46.296	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	92.600%		
35) Dibromofluoromethane	8.177	113	210985	49.569	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.140%		
50) Toluene-d8	10.577	98	806196	49.209	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.420%		
62) 4-Bromofluorobenzene	12.859	95	234844	40.587	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	81.180%#		
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	7.494	96	1646	0.322	ug/l	50
44) Trichloroethene	9.365	130	3660	0.787	ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
Data File : VN077935.D
Acq On : 26 May 2023 18:17
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 29 00:32:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

