

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
 Data File : VN077949.D
 Acq On : 27 May 2023 02:03
 Operator : JC\MD
 Sample : 02957-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-9-0A

Quant Time: May 29 01:41:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	374437	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	694176	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	649415	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	255919	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	254611	45.446	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.900%
35) Dibromofluoromethane	8.177	113	222877	49.237	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.480%
50) Toluene-d8	10.576	98	883855	50.728	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.460%
62) 4-Bromofluorobenzene	12.853	95	315820	51.323	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.640%
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	31127	17.659	ug/l	97
20) Methylene Chloride	5.283	84	26004	3.844	ug/l #	84
37) Ethyl Acetate	7.571	43	62204	10.541	ug/l	96
43) Isopropyl Acetate	8.700	43	297626	30.877	ug/l #	81
95) Naphthalene	15.647	128	13817333	789.802	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
Data File : VN077949.D
Acq On : 27 May 2023 02:03
Operator : JC\MD
Sample : 02957-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-9-0A

Quant Time: May 29 01:41:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

