

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
 Data File : VN077951.D
 Acq On : 27 May 2023 02:51
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: May 29 01:42:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	384463	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	689505	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	604032	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	203897	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	258646	44.962	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.920%
35) Dibromofluoromethane	8.177	113	220925	49.136	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.280%
50) Toluene-d8	10.577	98	855005	49.405	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.800%
62) 4-Bromofluorobenzene	12.859	95	254574	41.650	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.300%
Target Compounds						
95) Naphthalene	15.647	128	35926	2.577	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
Data File : VN077951.D
Acq On : 27 May 2023 02:51
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 29 01:42:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

