

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
 Data File : VN077952.D
 Acq On : 27 May 2023 03:15
 Operator : JC\MD
 Sample : 02970-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: May 29 01:42:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	372204	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	677772	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	577041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	176971	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	249349	44.773	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.540%
35) Dibromofluoromethane	8.177	113	222941	50.443	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.880%
50) Toluene-d8	10.571	98	822174	48.330	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.660%
62) 4-Bromofluorobenzene	12.859	95	259595	43.207	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.420%
Target Compounds						
95) Naphthalene	15.653	128	15732	1.300	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
Data File : VN077952.D
Acq On : 27 May 2023 03:15
Operator : JC\MD
Sample : 02970-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: May 29 01:42:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

