

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
 Data File : VN077954.D
 Acq On : 27 May 2023 04:03
 Operator : JC\MD
 Sample : 02970-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: May 29 01:43:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	363577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	677376	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	582001	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	179984	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	251667	46.262	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.520%
35) Dibromofluoromethane	8.177	113	220078	49.824	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.640%
50) Toluene-d8	10.576	98	830100	48.824	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.640%
62) 4-Bromofluorobenzene	12.859	95	265834	44.271	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.540%
Target Compounds						
95) Naphthalene	15.641	128	6404	0.520	ug/l	Qvalue # 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
Data File : VN077954.D
Acq On : 27 May 2023 04:03
Operator : JC\MD
Sample : 02970-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: May 29 01:43:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

