

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
 Data File : VN077955.D
 Acq On : 27 May 2023 04:27
 Operator : JC\MD
 Sample : 02970-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: May 29 01:43:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	360155	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	668201	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	579636	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	179427	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	256143	47.532	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.060%
35) Dibromofluoromethane	8.177	113	216419	49.669	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.340%
50) Toluene-d8	10.577	98	829134	49.437	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.880%
62) 4-Bromofluorobenzene	12.859	95	263124	44.422	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
Data File : VN077955.D
Acq On : 27 May 2023 04:27
Operator : JC\MD
Sample : 02970-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: May 29 01:43:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

