

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
 Data File : VN077941.D
 Acq On : 26 May 2023 22:41
 Operator : JC\MD
 Sample : VN0526WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0526WBL02

Quant Time: May 29 01:30:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	377269	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	674283	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	589895	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	185294	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	255485	45.259	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.520%
35) Dibromofluoromethane	8.177	113	217498	49.466	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.940%
50) Toluene-d8	10.577	98	832002	49.161	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.320%
62) 4-Bromofluorobenzene	12.859	95	270457	45.248	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.500%


Target Compounds Qvalue


 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052623\
Data File : VN077941.D
Acq On : 26 May 2023 22:41
Operator : JC\MD
Sample : VN0526WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0526WBL02

Quant Time: May 29 01:30:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.236 min Scan# 110
Delta R.T. 0.006 min
Lab File: VN077941.D
Acq: 26 May 2023 22:41




Instrument :
MSVOA_N
ClientSampleId :
VN0526WBL02


Tgt Ion:168 Resp: 377269
Ion Ratio Lower Upper
168 100
99 61.1 53.0 79.6



#33
1,2-Dichloroethane-d4
Concen: 45.259 ug/l
RT: 8.588 min Scan# 1128
Delta R.T. 0.006 min
Lab File: VN077941.D
Acq: 26 May 2023 22:41

Tgt Ion: 65 Resp: 255485
Ion Ratio Lower Upper
65 100
67 52.2 0.0 106.6









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.106 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN077941.D
Acq: 26 May 2023 22:41




Instrument :
MSVOA_N
ClientSampleId :
VN0526WBL02


Tgt Ion:114 Resp: 674283
Ion Ratio Lower Upper
114 100
63 18.9 0.0 36.4
88 16.3 0.0 30.6



#35
Dibromofluoromethane
Concen: 49.466 ug/l
RT: 8.177 min Scan# 1058
Delta R.T. 0.006 min
Lab File: VN077941.D
Acq: 26 May 2023 22:41

Tgt Ion:113 Resp: 217498
Ion Ratio Lower Upper
113 100
111 105.1 82.6 123.8
192 16.3 12.7 19.1









#50
Toluene-d8
Concen: 49.161 ug/l
RT: 10.577 min Scan# 1465
Delta R.T. 0.006 min
Lab File: VN077941.D
Acq: 26 May 2023 22:41




Instrument :
MSVOA_N
ClientSampleId :
VN0526WBL02


Tgt Ion: 98 Resp: 832002
Ion Ratio Lower Upper
98 100
100 63.8 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 45.248 ug/l
RT: 12.859 min Scan# 1854
Delta R.T. 0.006 min
Lab File: VN077941.D
Acq: 26 May 2023 22:41

Tgt Ion: 95 Resp: 270457
Ion Ratio Lower Upper
95 100
174 72.4 0.0 143.8
176 70.2 0.0 135.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 11
Delta R.T. 0.001 min
Lab File: VN077941.D
Acq: 26 May 2023 22:41

Instrument :
MSVOA_N
ClientSampleId :
VN0526WBL02

Tgt Ion:117 Resp: 589895
Ion Ratio Lower Upper
117 100
82 56.4 44.0 66.0
119 30.8 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.800 min Scan# 2014
Delta R.T. -0.000 min
Lab File: VN077941.D
Acq: 26 May 2023 22:41

Tgt Ion:152 Resp: 185294
Ion Ratio Lower Upper
152 100
115 63.2 32.3 96.8
150 161.5 0.0 355.8

