

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052819\
 Data File : VN055852.D
 Acq On : 28 May 2019 14:56
 Operator : JC/SP
 Sample : K3037-02DL 100X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW6-R1-2DL

Quant Time: May 28 17:39:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	407773	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	584798	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	533544	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	185243	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	194518	46.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.38%	
35) Dibromofluoromethane	7.59	113	175527	47.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.94%	
50) Toluene-d8	10.09	98	697032	49.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.30%	
62) 4-Bromofluorobenzene	12.40	95	223153	45.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.78%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.18	62	45919	11.197	ug/l	99
11) Tert butyl alcohol	4.74	59	5535	6.879	ug/l #	74
12) 1,1-Dichloroethene	3.72	96	1126	0.325	ug/l	95
16) Acetone	3.82	43	1158	0.626	ug/l #	88
18) Methyl Acetate	4.34	43	2986	Below Cal	#	80
27) cis-1,2-Dichloroethene	6.83	96	255817	58.852	ug/l	89
31) Cyclohexane	7.67	56	5647	0.893	ug/l #	1
36) 1,1-Dichloropropene	7.66	75	22945	4.502	ug/l #	47
38) Carbon Tetrachloride	7.66	117	33828	6.372	ug/l #	14
43) Isopropyl Acetate	8.32	43	8654	1.046	ug/l #	53
44) Trichloroethene	8.83	130	19827	4.630	ug/l	99
51) 4-Methyl-2-Pentanone	9.98	43	1331	0.272	ug/l #	35
60) Dibromochloromethane	10.63	129	21699	4.813	ug/l #	11
64) Tetrachloroethene	10.63	164	26051	5.789	ug/l	97
73) Isopropylbenzene	12.25	105	386	Below Cal	#	48
74) N-amyl acetate	12.07	43	217	Below Cal	#	43
76) 1,2,3-Trichloropropane	12.40	75	104326	31.545	ug/l #	100
79) 2-Chlorotoluene	12.67	91	62	Below Cal	#	39
80) 1,3,5-Trimethylbenzene	12.73	105	229	Below Cal	#	28
81) trans-1,4-Dichloro-2-buten	12.40	75	104326	75.775	ug/l #	15
84) 1,2,4-Trimethylbenzene	13.04	105	402	Below Cal	#	30
85) sec-Butylbenzene	13.17	105	369	Below Cal	#	55
93) 1,2,4-Trichlorobenzene	14.91	180	179	3.683	ug/l #	31
94) Hexachlorobutadiene	15.00	225	29	Below Cal	#	19
95) Naphthalene	15.13	128	1945	0.227	ug/l #	74


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052819\
Data File : VN055852.D
Acq On : 28 May 2019 14:56
Operator : JC/SP
Sample : K3037-02DL 100X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW6-R1-2DL

Quant Time: May 28 17:39:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.66 min Scan# 1832

Delta R.T. 0.00 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56

Instrument :

MSVOA_N

ClientSampleId :


MW6-R1-2DL

Tot Ion:168 Resp: 407773

Ion Ratio Lower Upper

168 100

99 46.2 37.2 55.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.66

150000

100000



50000

0

Time-->

7.60

7.70



#4

Vinyl Chloride

Concen: 11.197 ug/l

RT: 2.18 min Scan# 128

Delta R.T. 0.00 min

Lab File: VN055852.D


Acq: 28 May 2019 14:56

Tgt Ion: 62 Resp: 45919

Ion Ratio Lower Upper

62 100

64 31.3 25.6 38.4



Abundance Ion 61.90 (61.60 to 62.60): VNO

Ion 63.90 (63.60 to 64.60): VNO

2.18

30000

20000

10000

0


Time-->


2.10

2.15

2.20

2.25









#11
Tert butvl alcohol
Concen: 6.879 ug/l
RT: 4.74 min Scan# 922
Delta R.T. -0.07 min
Lab File: VN055852.D
Acq: 28 May 2019 14:56




Instrument :
MSVOA_N
ClientSampleId :
MW6-R1-2DL


Tot Ion: 59 Resp: 5535
Ion Ratio Lower Upper
59 100
57 0.0 7.8 11.6#



#12
1,1-Dichloroethene
Concen: 0.325 ug/l
RT: 3.72 min Scan# 606
Delta R.T. 0.00 min
Lab File: VN055852.D
Acq: 28 May 2019 14:56

Tgt Ion: 96 Resp: 1126
Ion Ratio Lower Upper
96 100
61 155.5 123.2 184.8
98 75.2 50.4 75.6





#16

Acetone

Concen: 0.626 ug/l

RT: 3.82 min Scan# 637

Delta R.T. 0.00 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56

Instrument :

MSVOA_N

ClientSampleId :





MW6-R1-2DL

Tot Ion: 43 Resp: 1158

Ion Ratio Lower Upper

43 100

58 26.5 26.6 39.8#



#18

Methyl Acetate

Concen: Below Cal

RT: 4.34 min Scan# 797

Delta R.T. 0.01 min

Lab File: VN055852.D




Acq: 28 May 2019 14:56


Tgt Ion: 43 Resp: 2986

Ion Ratio Lower Upper

43 100

74 14.0 19.3 28.9#





#27

cis-1,2-Dichloroethene

Concen: 58.852 ug/l

RT: 6.83 min Scan# 1572

Delta R.T. 0.00 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56

Instrument :

MSVOA_N

ClientSampled :

MW6-R1-2DL


Tot Ion: 96 Resp: 255817

Ion Ratio Lower Upper

96 100

61 118.9 0.0 275.8




98 64.6 0.0 128.4



Abundance Ion 95.90 (95.60 to 96.60): VN0

Ion 60.90 (60.60 to 61.60): VN0

Ion 97.90 (97.60 to 98.60): VN0



#31

Cyclohexane

Concen: 0.893 ug/l

RT: 7.67 min Scan# 1834

Delta R.T. 0.02 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56


Tgt Ion: 56 Resp: 5647

Ion Ratio Lower Upper

56 100

69 197.9 25.6 38.4#



84 147.9 67.4 101.0#




Abundance Ion 56.00 (55.70 to 56.70): VN0

Ion 69.00 (68.70 to 69.70): VN0

Ion 84.00 (83.70 to 84.70): VN0









#33
 1,2-Dichloroethane-d4
 Concen: 46.191 ug/l
 RT: 8.03 min Scan# 1945
 Delta R.T. 0.00 min
 Lab File: VN055852.D
 Acq: 28 May 2019 14:56




Instrument :
 MSVOA_N
 ClientSampleId :
 MW6-R1-2DL


Tot Ion: 65 Resp: 194518
 Ion Ratio Lower Upper
 65 100
 67 53.7 0.0 109.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN055852.D
 Acq: 28 May 2019 14:56

Tgt Ion: 114 Resp: 584798
 Ion Ratio Lower Upper
 114 100
 63 18.4 0.0 39.2
 88 14.9 0.0 31.0









#35
 Dibromofluoromethane
 Concen: 47.474 ug/l
 RT: 7.59 min Scan# 1809
 Delta R.T. 0.00 min
 Lab File: VN055852.D
 Acq: 28 May 2019 14:56




Instrument :
 MSVOA_N
 ClientSampleId :
 MW6-R1-2DL


Tot Ion:113 Resp: 175527
 Ion Ratio Lower Upper
 113 100
 111 102.6 81.9 122.9
 192 24.6 19.5 29.3



#36
 1,1-Dichloropropene
 Concen: 4.502 ug/l
 RT: 7.66 min Scan# 1831
 Delta R.T. -0.13 min
 Lab File: VN055852.D
 Acq: 28 May 2019 14:56

Tgt Ion: 75 Resp: 22945
 Ion Ratio Lower Upper
 75 100
 110 8.4 19.0 57.0#
 77 0.0 25.1 37.7#









#38
Carbon Tetrachloride
Concen: 6.372 ug/l
RT: 7.66 min Scan# 1832
Delta R.T. -0.11 min
Lab File: VN055852.D
Acq: 28 May 2019 14:56

Instrument :
MSVOA_N
ClientSampleId :
MW6-R1-2DL

Tot Ion:117 Resp: 33828
Ion Ratio Lower Upper
117 100
119 4.3 79.0 118.6#
121 0.0 24.7 37.1#




Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V






#43
Isopropyl Acetate
Concen: 1.046 ug/l
RT: 8.32 min Scan# 2037
Delta R.T. 0.16 min
Lab File: VN055852.D
Acq: 28 May 2019 14:56

Tgt Ion: 43 Resp: 8654
Ion Ratio Lower Upper
43 100
61 0.0 22.2 33.4#
87 0.0 11.4 17.2#



Abundance Ion 43.00 (42.70 to 43.70): V
Ion 61.00 (60.70 to 61.70): V
Ion 87.00 (86.70 to 87.70): V





#44

Trichloroethene

Concen: 4.630 ug/l

RT: 8.83 min Scan# 2196

Delta R.T. 0.00 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56

Instrument :

MSVOA_N

ClientSampled :


MW6-R1-2DL

Tot Ion:130 Resp: 19827




Ion Ratio Lower Upper

130 100

95 92.6 0.0 188.0



Abundance Ion 129.80 (129.50 to 130.50): VN055746.D (-2180) (-)



#50

Toluene-d8

Concen: 49.148 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN055852.D


Acq: 28 May 2019 14:56

Tgt Ion: 98 Resp: 697032



Ion Ratio Lower Upper


98 100

100 64.6 51.9 77.9



Abundance Ion 98.00 (97.70 to 98.70): VN055746.D (-2570) (-)





#51

4-Methyl-2-Pentanone

Concen: 0.272 ug/l

RT: 9.98 min Scan# 2552

Delta R.T. -0.01 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56

Instrument :

MSVOA_N

ClientSampled :


MW6-R1-2DL

Tot Ion: 43 Resp: 1331

Ion Ratio Lower Upper

43 100

58 0.0 32.1 48.1#



Abundance Ion 43.00 (42.70 to 43.70): VN055746.D (-2537) (-)

Ion 58.00 (57.70 to 58.70): VN055746.D (-2537) (-)

1500

1000

500

0



9.95

9.98

10.00

10.05

Time-->



#60

Dibromochloromethane

Concen: 4.813 ug/l

RT: 10.63 min Scan# 2755

Delta R.T. -0.27 min

Lab File: VN055852.D


Acq: 28 May 2019 14:56

Tgt Ion: 129 Resp: 21699

Ion Ratio Lower Upper

129 100

127 0.0 38.1 114.3#



Abundance Ion 128.80 (128.50 to 129.50): VN055746.D (-2824) (-)

Ion 126.80 (126.50 to 127.50): VN055746.D (-2824) (-)

15000

10000

5000

0


10.60


10.63

10.65

10.70

Time-->









#62
4-Bromofluorobenzene
Concen: 45.888 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.00 min
Lab File: VN055852.D
Acq: 28 May 2019 14:56




Instrument :
MSVOA_N
ClientSampleId :
MW6-R1-2DL


Tot Ion: 95 Resp: 223153
Ion Ratio Lower Upper
95 100
174 97.9 0.0 191.0
176 94.1 0.0 185.2



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN055852.D
Acq: 28 May 2019 14:56

Tgt Ion: 117 Resp: 533544
Ion Ratio Lower Upper
117 100
82 52.2 41.8 62.8
119 31.8 25.4 38.0





#64

Tetrachloroethene

Concen: 5.789 ug/l

RT: 10.63 min Scan# 2755

Delta R.T. 0.00 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56

Instrument :

MSVOA_N

ClientSampleId :

MW6-R1-2DL

Tot Ion:164 Resp: 26051


Ion Ratio Lower Upper

164 100

166 134.3 103.0 154.6

129 90.4 70.6 105.8

131 86.7 68.6 102.8






Abundance Ion 163.80 (163.50 to 164.50): V

Ion 165.80 (165.50 to 166.50): V

Ion 128.80 (128.50 to 129.50): V

Ion 130.80 (130.50 to 131.50): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3598

Delta R.T. 0.00 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56


Tgt Ion:152 Resp: 185243

Ion Ratio Lower Upper

152 100

115 52.9 27.0 81.0



150 156.7 0.0 343.2




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: Below Cal

RT: 12.25 min Scan# 3258

Delta R.T. 0.00 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56

Instrument :

MSVOA_N

ClientSampleId :


MW6-R1-2DL

Tot Ion:105 Resp: 386

Ion Ratio Lower Upper

105 100

120 0.0 13.6 40.7#



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

300

250

200

150



100

50

0

12.22 12.24 12.26 12.28

Time-->



#74

N-ethyl acetate

Concen: Below Cal

RT: 12.07 min Scan# 3203

Delta R.T. 0.00 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56

Tgt Ion: 43 Resp: 217


Ion Ratio Lower Upper

43 100

70 0.0 34.0 51.0#

55 0.0 18.6 28.0#

61 0.0 20.0 30.0#



Abundance Ion 43.00 (42.70 to 43.70): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 61.00 (60.70 to 61.70): V

300


200


100

0

12.06 12.08

Time-->





#76

1,2,3-Trichloropropane

Concen: 31.545 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.15 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56

Instrument :

MSVOA_N

ClientSampled :





MW6-R1-2DL

Tot Ion: 75 Resp: 104326

Ion Ratio Lower Upper

75 100

77 1.2 0.0 0.0#



#79

2-Chlorotoluene

Concen: Below Cal

RT: 12.67 min Scan# 3389

Delta R.T. -0.00 min

Lab File: VN055852.D




Acq: 28 May 2019 14:56


Tgt Ion: 91 Resp: 62

Ion Ratio Lower Upper

91 100

126 0.0 18.1 54.4#





#80

1,3,5-Trimethylbenzene

Concen: Below Cal

RT: 12.73 min Scan# 3409

Delta R.T. 0.00 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56

Instrument :

MSVOA_N

ClientSampled :


MW6-R1-2DL

Tot Ion:105 Resp: 229

Ion Ratio Lower Upper

105 100

120 0.0 24.7 74.1#



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

12.73



300

200

100

0

Time-->



#81

trans-1,4-Dichloro-2-butene

Concen: 75.775 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.10 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56


Tgt Ion: 75 Resp: 104326

Ion Ratio Lower Upper

75 100

53 0.0 68.1 102.1#

89 0.0 36.9 55.3#



Abundance Ion 74.90 (74.60 to 75.60): VNO

Ion 53.00 (52.70 to 53.70): VNO

Ion 88.90 (88.60 to 89.60): VNO

80000


60000


40000

20000

0

Time-->









#84
 1,2,4-Trimethylbenzene
 Concen: Below Cal
 RT: 13.04 min Scan# 3504
 Delta R.T. 0.00 min
 Lab File: VN055852.D
 Acq: 28 May 2019 14:56




Instrument :
 MSVOA_N
 ClientSampled :
 MW6-R1-2DL


Tot Ion:105 Resp: 402
 Ion Ratio Lower Upper
 105 100
 120 0.0 23.0 69.0#



#85
 sec-Butylbenzene
 Concen: Below Cal
 RT: 13.17 min Scan# 3546
 Delta R.T. 0.00 min
 Lab File: VN055852.D
 Acq: 28 May 2019 14:56

Tgt Ion:105 Resp: 369
 Ion Ratio Lower Upper
 105 100
 134 0.0 10.4 31.2#





#93

1,2,4-Trichlorobenzene

Concen: 3.683 ug/l

RT: 14.91 min Scan# 4086

Delta R.T. 0.00 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56

Instrument :

MSVOA_N

ClientSampled :

MW6-R1-2DL


Tot Ion:180 Resp: 179

Ion Ratio Lower Upper

180 100

182 168.7 48.3 144.9#


145 0.0 14.0 41.9#





Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.80 (144.50 to 145.50): V



Time-->



#94

Hexachlorobutadiene

Concen: Below Cal

RT: 15.00 min Scan# 4115

Delta R.T. -0.00 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56


Tgt Ion:225 Resp: 29

Ion Ratio Lower Upper

225 100

223 0.0 30.9 92.7#


227 0.0 31.6 94.7#




Abundance Ion 224.70 (224.40 to 225.40): V


Ion 222.70 (222.40 to 223.40): V

Ion 226.70 (226.40 to 227.40): V



Time-->





#95

Naphthalene

Concen: 0.227 ug/l

RT: 15.13 min Scan# 4155

Delta R.T. 0.00 min

Lab File: VN055852.D

Acq: 28 May 2019 14:56

Instrument :

MSVOA_N

ClientSampleId :

MW6-R1-2DL


Tot Ion:128 Resp: 1945

Ion Ratio Lower Upper

128 100

127 3.2 10.3 15.5#

129 0.0 8.7 13.1#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

