

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN052819\
 Data File : VN055839.D
 Acq On : 28 May 2019 9:30
 Operator : JC/SP
 Sample : VN0528MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0528MBL01

Quant Time: May 28 18:05:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	401808	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	564798	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	506661	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	164034	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	188417	45.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.82%	
35) Dibromofluoromethane	7.59	113	171590	48.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.10%	
50) Toluene-d8	10.09	98	679349	49.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.20%	
62) 4-Bromofluorobenzene	12.40	95	207944	44.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.54%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052819\
Data File : VN055839.D
Acq On : 28 May 2019 9:30
Operator : JC/SP
Sample : VN0528MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0528MBL01

Quant Time: May 28 18:05:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

