

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052819\
 Data File : VN055844.D
 Acq On : 28 May 2019 11:36
 Operator : JC/SP
 Sample : K3040-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 0444-FIELD-BLANK-1

Quant Time: May 28 18:33:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	431202	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	604775	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	547357	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	182645	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	201398	45.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.46%	
35) Dibromofluoromethane	7.59	113	183285	47.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.88%	
50) Toluene-d8	10.09	98	716821	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
62) 4-Bromofluorobenzene	12.40	95	223889	44.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.04%	
Target Compounds						
16) Acetone	3.82	43	11110	5.676	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052819\
Data File : VN055844.D
Acq On : 28 May 2019 11:36
Operator : JC/SP
Sample : K3040-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
0444-FIELD-BLANK-1

Quant Time: May 28 18:33:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

