

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052819\
 Data File : VN055853.D
 Acq On : 28 May 2019 15:21
 Operator : JC/SP
 Sample : K3037-01DL 100X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW6-R1-1DL

Quant Time: May 29 05:53:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	430061	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	614900	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	557929	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	191583	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	201082	45.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.56%	
35) Dibromofluoromethane	7.59	113	181029	46.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.14%	
50) Toluene-d8	10.09	98	726638	48.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.46%	
62) 4-Bromofluorobenzene	12.40	95	232096	45.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.78%	
Target Compounds						
4) Vinyl Chloride	2.18	62	38069	8.801	ug/l	97
27) cis-1,2-Dichloroethene	6.83	96	221366	48.287	ug/l	89
44) Trichloroethene	8.83	130	8234	1.829	ug/l	98
64) Tetrachloroethene	10.63	164	7290	1.549	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052819\
Data File : VN055853.D
Acq On : 28 May 2019 15:21
Operator : JC/SP
Sample : K3037-01DL 100X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW6-R1-1DL

Quant Time: May 29 05:53:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

