

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052821\
 Data File : VN067180.D
 Acq On : 28 May 2021 13:58
 Operator : JC/MD
 Sample : VN0528WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0528WBL01

Quant Time: May 29 02:19:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 21 17:24:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.578	168	241164	50.00	ug/l	#-0.02
34) 1,4-Difluorobenzene	8.509	114	432731	50.00	ug/l	-0.02
63) Chlorobenzene-d5	11.344	117	362173	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.286	152	142817	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.943	65	170408	45.53	ug/l	-0.02
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.06%
35) Dibromofluoromethane	7.500	113	133180	50.15	ug/l	-0.02
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.30%
50) Toluene-d8	10.022	98	516953	50.58	ug/l	-0.01
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.16%
62) 4-Bromofluorobenzene	12.344	95	171404	49.06	ug/l	-0.01
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.12%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052821\
Data File : VN067180.D
Acq On : 28 May 2021 13:58
Operator : JC/MD
Sample : VN0528WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0528WBL01

Quant Time: May 29 02:19:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
Quant Title : SW846 8260
QLast Update : Fri May 21 17:24:08 2021
Response via : Initial Calibration

