

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052821\
 Data File : VN067183.D
 Acq On : 28 May 2021 15:15
 Operator : JC/MD
 Sample : PB136697TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136697TB

Manual Integrations
 APPROVED

MMDadoda
 6/1/2021 12:13:34 PM

Quant Time: May 29 02:23:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 21 17:24:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.581	168	251932	50.00	ug/l	#-0.02
34) 1,4-Difluorobenzene	8.509	114	459527	50.00	ug/l	-0.02
63) Chlorobenzene-d5	11.344	117	411220	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.285	152	148849	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.943	65	184049	47.07	ug/l	-0.02
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.14%		
35) Dibromofluoromethane	7.503	113	142727	50.61	ug/l	-0.02
Spiked Amount 50.000	Range 69 - 133		Recovery =	101.22%		
50) Toluene-d8	10.022	98	568177	52.35	ug/l	-0.01
Spiked Amount 50.000	Range 65 - 126		Recovery =	104.70%		
62) 4-Bromofluorobenzene	12.341	95	194327	52.38	ug/l	-0.01
Spiked Amount 50.000	Range 58 - 135		Recovery =	104.76%		
Target Compounds						Qvalue
16) Acetone	3.746	43	24651	15.34	ug/l	92
20) Methylene Chloride	4.462	84	23400	5.92	ug/l #	86
43) Isopropyl Acetate	8.085	43	38178	4.30	ug/l #	84
95) Naphthalene	15.082	128	10295m	5.12	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052821\
Data File : VN067183.D
Acq On : 28 May 2021 15:15
Operator : JC/MD
Sample : PB136697TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136697TB

Manual Integrations
APPROVED

MMDadoda
6/1/2021 12:13:34 PM

Quant Time: May 29 02:23:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
Quant Title : SW846 8260
QLast Update : Fri May 21 17:24:08 2021
Response via : Initial Calibration

