

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052821\
 Data File : VN067184.D
 Acq On : 28 May 2021 15:39
 Operator : JC/MD
 Sample : PB136697ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136697ZHE#01

Quant Time: May 29 02:23:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 21 17:24:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.581	168	251335	50.00	ug/l	#-0.02
34) 1,4-Difluorobenzene	8.509	114	452322	50.00	ug/l	-0.02
63) Chlorobenzene-d5	11.344	117	392345	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.285	152	138270	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.943	65	184551	47.31	ug/l	-0.02
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.62%		
35) Dibromofluoromethane	7.503	113	141689	51.04	ug/l	-0.02
Spiked Amount 50.000	Range 69 - 133		Recovery =	102.08%		
50) Toluene-d8	10.021	98	557440	52.18	ug/l	-0.01
Spiked Amount 50.000	Range 65 - 126		Recovery =	104.36%		
62) 4-Bromofluorobenzene	12.341	95	184026	50.39	ug/l	-0.01
Spiked Amount 50.000	Range 58 - 135		Recovery =	100.78%		
Target Compounds						
16) Acetone	3.740	43	78774	49.13	ug/l	95
20) Methylene Chloride	4.459	84	26248	6.66	ug/l #	87
43) Isopropyl Acetate	8.088	43	33826	3.87	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052821\
Data File : VN067184.D
Acq On : 28 May 2021 15:39
Operator : JC/MD
Sample : PB136697ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136697ZHE#01

Quant Time: May 29 02:23:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
Quant Title : SW846 8260
QLast Update : Fri May 21 17:24:08 2021
Response via : Initial Calibration

