

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052821\
 Data File : VN067185.D
 Acq On : 28 May 2021 16:02
 Operator : JC/MD
 Sample : PB136697ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136697ZHE#02

Quant Time: May 29 02:24:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 21 17:24:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.583	168	250090	50.00	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.509	114	459640	50.00	ug/l	-0.02
63) Chlorobenzene-d5	11.344	117	399464	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.285	152	143352	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.943	65	182690	47.07	ug/l	-0.02
Spiked Amount 50.000	Range 61 - 141		Recovery = 94.14%			
35) Dibromofluoromethane	7.506	113	143024	50.70	ug/l	-0.01
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.40%			
50) Toluene-d8	10.021	98	558891	51.48	ug/l	-0.01
Spiked Amount 50.000	Range 65 - 126		Recovery = 102.96%			
62) 4-Bromofluorobenzene	12.341	95	188753	50.86	ug/l	-0.01
Spiked Amount 50.000	Range 58 - 135		Recovery = 101.72%			
Target Compounds						
16) Acetone	3.743	43	78913	49.46	ug/l	92
20) Methylene Chloride	4.462	84	26242	6.69	ug/l	89
43) Isopropyl Acetate	8.085	43	36628	4.12	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052821\
Data File : VN067185.D
Acq On : 28 May 2021 16:02
Operator : JC/MD
Sample : PB136697ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136697ZHE#02

Quant Time: May 29 02:24:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
Quant Title : SW846 8260
QLast Update : Fri May 21 17:24:08 2021
Response via : Initial Calibration

