

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052821\
 Data File : VN067189.D
 Acq On : 28 May 2021 17:35
 Operator : JC/MD
 Sample : PB136697ZHE#06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136697ZHE#06

Quant Time: May 29 02:25:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 21 17:24:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.583	168	237890	50.00	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.514	114	428252	50.00	ug/l	-0.01
63) Chlorobenzene-d5	11.349	117	379264	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.291	152	139689	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	172310	46.67	ug/l	-0.01
Spiked Amount 50.000	Range 61 - 141		Recovery =	93.34%		
35) Dibromofluoromethane	7.506	113	134963	51.35	ug/l	-0.01
Spiked Amount 50.000	Range 69 - 133		Recovery =	102.70%		
50) Toluene-d8	10.027	98	528982	52.30	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	104.60%		
62) 4-Bromofluorobenzene	12.349	95	172682	49.94	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	99.88%		
Target Compounds						
16) Acetone	3.743	43	73812	48.64	ug/l	96
20) Methylene Chloride	4.459	84	24098	6.46	ug/l	91
43) Isopropyl Acetate	8.088	43	36276	4.38	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052821\
Data File : VN067189.D
Acq On : 28 May 2021 17:35
Operator : JC/MD
Sample : PB136697ZHE#06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136697ZHE#06

Quant Time: May 29 02:25:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
Quant Title : SW846 8260
QLast Update : Fri May 21 17:24:08 2021
Response via : Initial Calibration

