

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052821\
 Data File : VN067193.D
 Acq On : 28 May 2021 19:09
 Operator : JC/MD
 Sample : PB136697ZHE#10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136697ZHE#10

Quant Time: May 29 02:26:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 21 17:24:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.583	168	246245	50.00	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.511	114	443983	50.00	ug/l	-0.01
63) Chlorobenzene-d5	11.349	117	394895	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.288	152	144096	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	178842	46.80	ug/l	-0.01
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.60%
35) Dibromofluoromethane	7.506	113	137496	50.46	ug/l	-0.01
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.92%
50) Toluene-d8	10.027	98	538837	51.39	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.78%
62) 4-Bromofluorobenzene	12.347	95	186998	52.16	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.32%
Target Compounds						
16) Acetone	3.740	43	21916	13.95	ug/l	98
18) Methyl Acetate	4.247	43	12282	2.80	ug/l #	61
20) Methylene Chloride	4.459	84	21584	5.59	ug/l	89
43) Isopropyl Acetate	8.090	43	33088	3.85	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052821\
Data File : VN067193.D
Acq On : 28 May 2021 19:09
Operator : JC/MD
Sample : PB136697ZHE#10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136697ZHE#10

Quant Time: May 29 02:26:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
Quant Title : SW846 8260
QLast Update : Fri May 21 17:24:08 2021
Response via : Initial Calibration

