

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052821\
 Data File : VN067196.D
 Acq On : 28 May 2021 23:06
 Operator : JC/MD
 Sample : M2551-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JCSY-SP2-VOC

Manual Integrations
 APPROVED

MMDadoda
 6/1/2021 12:14:05 PM

Quant Time: May 29 02:27:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 21 17:24:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.581	168	254193	50.00	ug/l	#-0.02
34) 1,4-Difluorobenzene	8.511	114	467376	50.00	ug/l	-0.01
63) Chlorobenzene-d5	11.346	117	406361	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.288	152	151912	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.943	65	185318	46.98	ug/l	-0.02
Spiked Amount 50.000	Range 61 - 141		Recovery =	93.96%		
35) Dibromofluoromethane	7.503	113	144536	50.39	ug/l	-0.02
Spiked Amount 50.000	Range 69 - 133		Recovery =	100.78%		
50) Toluene-d8	10.024	98	566674	51.34	ug/l	-0.01
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.68%		
62) 4-Bromofluorobenzene	12.344	95	188943	50.07	ug/l	-0.01
Spiked Amount 50.000	Range 58 - 135		Recovery =	100.14%		
Target Compounds						Qvalue
16) Acetone	3.746	43	30314	18.69	ug/l	93
18) Methyl Acetate	4.247	43	9010	1.99	ug/l #	77
20) Methylene Chloride	4.456	84	9714m	2.44	ug/l	
43) Isopropyl Acetate	8.085	43	21259	2.35	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052821\
Data File : VN067196.D
Acq On : 28 May 2021 23:06
Operator : JC/MD
Sample : M2551-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JCSY-SP2-VOC

Manual Integrations
APPROVED

MMDadoda
6/1/2021 12:14:05 PM

Quant Time: May 29 02:27:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N052121W.M
Quant Title : SW846 8260
QLast Update : Fri May 21 17:24:08 2021
Response via : Initial Calibration

