

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052824\
 Data File : VN082339.D
 Acq On : 28 May 2024 11:52
 Operator : JC\MD
 Sample : VN0528WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0528WBL01

Quant Time: May 29 06:46:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	235955	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	405150	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	376381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	164800	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	179610	48.207	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.420%
35) Dibromofluoromethane	8.171	113	129137	51.814	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.620%
50) Toluene-d8	10.565	98	507470	53.109	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.220%
62) 4-Bromofluorobenzene	12.847	95	182988	53.654	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	107.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052824\
Data File : VN082339.D
Acq On : 28 May 2024 11:52
Operator : JC\MD
Sample : VN0528WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0528WBL01

Quant Time: May 29 06:46:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

