

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052824\
 Data File : VN082346.D
 Acq On : 28 May 2024 14:53
 Operator : JC\MD
 Sample : P2606-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: May 29 06:51:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	205065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	357663	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	328783	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	136329	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	160368	49.526	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.060%
35) Dibromofluoromethane	8.171	113	114889	52.218	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.440%
50) Toluene-d8	10.571	98	443984	52.634	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.260%
62) 4-Bromofluorobenzene	12.853	95	156610	52.017	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	104.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052824\
Data File : VN082346.D
Acq On : 28 May 2024 14:53
Operator : JC\MD
Sample : P2606-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: May 29 06:51:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

