

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052824\
 Data File : VN082348.D
 Acq On : 28 May 2024 15:42
 Operator : JC\MD
 Sample : P2606-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: May 29 06:51:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	213010	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	371923	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	355183	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	150289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	165312	49.149	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.300%
35) Dibromofluoromethane	8.171	113	119199	52.100	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.200%
50) Toluene-d8	10.571	98	462048	52.675	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.360%
62) 4-Bromofluorobenzene	12.853	95	171273	54.706	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	109.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052824\
Data File : VN082348.D
Acq On : 28 May 2024 15:42
Operator : JC\MD
Sample : P2606-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: May 29 06:51:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

