

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052824\
 Data File : VN082356.D
 Acq On : 28 May 2024 18:56
 Operator : JC\MD
 Sample : P2609-03 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 420

Quant Time: May 29 06:57:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	196892	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	342918	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	325151	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	145618	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	154070	49.557	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.120%
35) Dibromofluoromethane	8.171	113	109825	52.063	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.120%
50) Toluene-d8	10.570	98	429503	53.107	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.220%
62) 4-Bromofluorobenzene	12.853	95	155893	54.005	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	108.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052824\
Data File : VN082356.D
Acq On : 28 May 2024 18:56
Operator : JC\MD
Sample : P2609-03 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
420

Quant Time: May 29 06:57:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

