

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052824\
 Data File : VN082360.D
 Acq On : 28 May 2024 20:33
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 29 07:00:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.721	0.608	15.7	83	0.00
3 P	Chloromethane	0.817	0.600	26.6#	76	0.00
4 C	Vinyl Chloride	0.768	0.631	17.8#	83	0.00
5 T	Bromomethane	0.441	0.336	23.8	82	-0.02
6 T	Chloroethane	0.501	0.426	15.0	91	0.00
7 T	Trichlorofluoromethane	1.006	0.979	2.7	98	-0.01
8 T	Diethyl Ether	0.358	0.378	-5.6	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.588	0.547	7.0	97	0.00
10 T	Methyl Iodide	0.591	0.568	3.9	93	0.00
11 T	Tert butyl alcohol	0.127	0.133	-4.7	106	0.00
12 CM	1,1-Dichloroethene	0.546	0.513	6.0#	94	0.00
13 T	Acrolein	0.094	0.083	11.7	85	0.00
14 T	Allyl chloride	1.169	1.083	7.4	98	0.00
15 T	Acrylonitrile	0.328	0.324	1.2	98	0.00
16 T	Acetone	0.324	0.291	10.2	86	0.00
17 T	Carbon Disulfide	1.764	1.325	24.9	80	0.00
18 T	Methyl Acetate	0.807	0.816	-1.1	106	0.00
19 T	Methyl tert-butyl Ether	1.996	2.059	-3.2	102	0.00
20 T	Methylene Chloride	0.653	0.605	7.4	98	0.00
21 T	trans-1,2-Dichloroethene	0.595	0.549	7.7	96	0.00
22 T	Diisopropyl ether	2.441	2.268	7.1	92	0.00
23 T	Vinyl Acetate	1.925	1.830	4.9	91	0.00
24 P	1,1-Dichloroethane	1.258	1.161	7.7	95	0.00
25 T	2-Butanone	0.491	0.472	3.9	95	0.00
26 T	2,2-Dichloropropane	1.054	0.887	15.8	84	0.00
27 T	cis-1,2-Dichloroethene	0.702	0.683	2.7	98	0.00
28 T	Bromochloromethane	0.586	0.495	15.5	92	0.00
29 T	Tetrahydrofuran	0.315	0.311	1.3	96	0.00
30 C	Chloroform	1.202	1.195	0.6#	102	0.00
31 T	Cyclohexane	1.261	0.968	23.2	85	0.00
32 T	1,1,1-Trichloroethane	1.066	1.087	-2.0	102	0.00
33 S	1,2-Dichloroethane-d4	0.790	0.729	7.7	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.308	0.308	0.0	100	0.00
36 T	1,1-Dichloropropene	0.516	0.483	6.4	93	0.00
37 T	Ethyl Acetate	0.577	0.569	1.4	93	0.00
38 T	Carbon Tetrachloride	0.523	0.545	-4.2	102	0.00
39 T	Methylcyclohexane	0.565	0.506	10.4	85	0.00
40 TM	Benzene	1.538	1.492	3.0	96	0.00
41 T	Methacrylonitrile	0.340	0.327	3.8	96	0.00
42 TM	1,2-Dichloroethane	0.592	0.586	1.0	99	0.00
43 T	Isopropyl Acetate	1.023	1.059	-3.5	101	0.00
44 TM	Trichloroethene	0.349	0.343	1.7	99	0.00
45 C	1,2-Dichloropropane	0.397	0.382	3.8#	96	0.00
46 T	Dibromomethane	0.261	0.267	-2.3	102	0.00
47 T	Bromodichloromethane	0.550	0.570	-3.6	101	0.00
48 T	Methyl methacrylate	0.487	0.495	-1.6	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052824\
 Data File : VN082360.D
 Acq On : 28 May 2024 20:33
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 29 07:00:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	110	0.00
50 S	Toluene-d8	1.179	1.132	4.0	95	0.00
51 T	4-Methyl-2-Pentanone	0.581	0.606	-4.3	99	0.00
52 CM	Toluene	0.923	0.939	-1.7#	98	0.00
53 T	t-1,3-Dichloropropene	0.569	0.591	-3.9	97	0.00
54 T	cis-1,3-Dichloropropene	0.602	0.611	-1.5	96	0.00
55 T	1,1,2-Trichloroethane	0.335	0.353	-5.4	105	0.00
56 T	Ethyl methacrylate	0.570	0.637	-11.8	102	0.00
57 T	1,3-Dichloropropane	0.627	0.625	0.3	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.294	0.301	-2.4	95	0.00
59 T	2-Hexanone	0.426	0.452	-6.1	97	0.00
60 T	Dibromochloromethane	0.370	0.427	-15.4	106	0.00
61 T	1,2-Dibromoethane	0.345	0.364	-5.5	105	0.00
62 S	4-Bromofluorobenzene	0.421	0.425	-1.0	98	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
64 T	Tetrachloroethene	0.371	0.364	1.9	100	0.00
65 PM	Chlorobenzene	1.105	1.107	-0.2	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.364	0.396	-8.8	105	0.00
67 C	Ethyl Benzene	1.976	2.001	-1.3#	97	0.00
68 T	m/p-Xylenes	0.724	0.747	-3.2	97	0.00
69 T	o-Xylene	0.701	0.735	-4.9	99	0.00
70 T	Styrene	1.165	1.250	-7.3	99	0.00
71 P	Bromoform	0.261	0.323	-23.8	112	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.972	4.023	-1.3	100	0.00
74 T	N-amyl acetate	2.185	2.153	1.5	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.145	1.160	-1.3	105	0.00
76 T	1,2,3-Trichloropropane	1.052	1.111	-5.6	105	0.00
77 T	Bromobenzene	0.947	0.957	-1.1	104	0.00
78 T	n-propylbenzene	4.704	4.657	1.0	95	0.00
79 T	2-Chlorotoluene	2.954	2.865	3.0	98	0.00
80 T	1,3,5-Trimethylbenzene	3.212	3.259	-1.5	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.450	0.434	3.6	96	0.00
82 T	4-Chlorotoluene	2.888	2.832	1.9	98	0.00
83 T	tert-Butylbenzene	2.810	2.839	-1.0	99	0.00
84 T	1,2,4-Trimethylbenzene	3.286	3.347	-1.9	97	0.00
85 T	sec-Butylbenzene	3.895	3.935	-1.0	97	0.00
86 T	p-Isopropyltoluene	3.165	3.238	-2.3	96	0.00
87 T	1,3-Dichlorobenzene	1.715	1.729	-0.8	101	0.00
88 T	1,4-Dichlorobenzene	1.750	1.706	2.5	100	0.00
89 T	n-Butylbenzene	2.845	2.796	1.7	92	0.00
90 T	Hexachloroethane	0.586	0.609	-3.9	100	0.00
91 T	1,2-Dichlorobenzene	1.623	1.663	-2.5	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.235	0.260	-10.6	105	0.00
93 T	1,2,4-Trichlorobenzene	0.910	0.891	2.1	95	0.00
94 T	Hexachlorobutadiene	0.399	0.365	8.5	94	0.00
95 T	Naphthalene	2.852	3.182	-11.6	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052824\
Data File : VN082360.D
Acq On : 28 May 2024 20:33
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: May 29 07:00:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.911	0.898	1.4	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6