

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052824\
 Data File : VN082355.D
 Acq On : 28 May 2024 18:32
 Operator : JC\MD
 Sample : P2609-02 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 414

Quant Time: May 29 06:56:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	201164	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	351532	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	328840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	141889	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	160629	50.569	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.140%
35) Dibromofluoromethane	8.171	113	115262	53.301	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.600%
50) Toluene-d8	10.571	98	437074	52.718	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.440%
62) 4-Bromofluorobenzene	12.847	95	153799	51.974	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	103.940%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052824\
Data File : VN082355.D
Acq On : 28 May 2024 18:32
Operator : JC\MD
Sample : P2609-02 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
414

Quant Time: May 29 06:56:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN082355.D
Acq: 28 May 2024 18:32




Instrument :
MSVOA_N
ClientSampleId :
414


Tgt Ion:168 Resp: 201164
Ion Ratio Lower Upper
168 100
99 60.0 48.2 72.4



#33
1,2-Dichloroethane-d4
Concen: 50.569 ug/l
RT: 8.583 min Scan# 1127
Delta R.T. 0.006 min
Lab File: VN082355.D
Acq: 28 May 2024 18:32

Tgt Ion: 65 Resp: 160629
Ion Ratio Lower Upper
65 100
67 48.8 0.0 103.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 114

Delta R.T. 0.006 min

Lab File: VN082355.D

Acq: 28 May 2024 18:32

Instrument :

MSVOA_N

ClientSampleId :

414





Tgt Ion:114 Resp: 351532

Ion Ratio Lower Upper

114 100

63 22.9 0.0 41.0

88 16.2 0.0 31.4



#35

Dibromofluoromethane

Concen: 53.301 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. 0.006 min

Lab File: VN082355.D

Acq: 28 May 2024 18:32




Tgt Ion:113 Resp: 115262


Ion Ratio Lower Upper

113 100

111 103.6 83.4 125.2

192 18.6 14.0 21.0









#50
Toluene-d8
Concen: 52.718 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. 0.000 min
Lab File: VN082355.D
Acq: 28 May 2024 18:32




Instrument :
MSVOA_N
ClientSampleId :
414


Tgt Ion: 98 Resp: 437074
Ion Ratio Lower Upper
98 100
100 64.8 51.8 77.6



#62
4-Bromofluorobenzene
Concen: 51.974 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN082355.D
Acq: 28 May 2024 18:32

Tgt Ion: 95 Resp: 153799
Ion Ratio Lower Upper
95 100
174 77.9 0.0 160.6
176 74.6 0.0 152.8








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN082355.D
Acq: 28 May 2024 18:32

Instrument :
MSVOA_N
ClientSampleId :
414

Tgt Ion:117 Resp: 328840
Ion Ratio Lower Upper
117 100
82 57.4 44.6 66.8
119 30.3 25.9 38.9



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN082355.D
Acq: 28 May 2024 18:32

Tgt Ion:152 Resp: 141889
Ion Ratio Lower Upper
152 100
115 61.5 30.4 91.3
150 157.0 0.0 353.2

