

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052825\
 Data File : VN086792.D
 Acq On : 28 May 2025 11:59
 Operator : JC\MD
 Sample : VN0528WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0528WBL01

Quant Time: May 29 03:55:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 17 00:46:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	190621	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	373140	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	366675	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	154289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	129060	48.997	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.000%
35) Dibromofluoromethane	8.165	113	105784	47.603	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.200%
50) Toluene-d8	10.565	98	482064	52.347	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.700%
62) 4-Bromofluorobenzene	12.847	95	173575	53.381	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052825\
Data File : VN086792.D
Acq On : 28 May 2025 11:59
Operator : JC\MD
Sample : VN0528WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0528WBL01

Quant Time: May 29 03:55:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051625W.M
Quant Title : SW846 8260
QLast Update : Sat May 17 00:46:13 2025
Response via : Initial Calibration

