

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052919\
 Data File : VN055879.D
 Acq On : 29 May 2019 15:48
 Operator : JC/SP
 Sample : K3085-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TANK-3

Quant Time: May 30 07:20:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N052219W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 23 05:54:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	50439	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	264807	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	216632	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	101944	30.92	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	103.07%
60) 4-Bromofluorobenzene	12.40	95	81994	24.33	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	81.10%
63) Toluene-d8	10.09	98	316593	31.98	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	106.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052919\
Data File : VN055879.D
Acq On : 29 May 2019 15:48
Operator : JC/SP
Sample : K3085-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TANK-3

Quant Time: May 30 07:20:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N052219W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 23 05:54:54 2019
Response via : Initial Calibration

