

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052920\
 Data File : VN061565.D
 Acq On : 29 May 2020 14:25
 Operator : JC/MD
 Sample : L2803-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-5

Quant Time: May 30 06:34:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	278774	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	502071	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	482768	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	209076	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	176778	53.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.44%	
35) Dibromofluoromethane	7.55	113	139415	54.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.48%	
50) Toluene-d8	10.06	98	609033	51.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.20%	
62) 4-Bromofluorobenzene	12.38	95	220964	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052920\
Data File : VN061565.D
Acq On : 29 May 2020 14:25
Operator : JC/MD
Sample : L2803-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-5

Quant Time: May 30 06:34:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

