

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052920\
 Data File : VN061571.D
 Acq On : 29 May 2020 16:46
 Operator : JC/MD
 Sample : PB129155ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129155ZHE#03

Quant Time: May 30 06:47:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	271793	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	490095	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	498319	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	227005	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	172269	53.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.40%	
35) Dibromofluoromethane	7.55	113	128898	51.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.70%	
50) Toluene-d8	10.06	98	616092	53.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.94%	
62) 4-Bromofluorobenzene	12.38	95	232092	55.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.44%	
Target Compounds						
16) Acetone	3.78	43	11966	7.390	ug/l	90
43) Isopropyl Acetate	8.13	43	67292	10.703	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052920\
Data File : VN061571.D
Acq On : 29 May 2020 16:46
Operator : JC/MD
Sample : PB129155ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129155ZHE#03

Quant Time: May 30 06:47:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

