

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN052920\
 Data File : VN061576.D
 Acq On : 29 May 2020 18:43
 Operator : JC/MD
 Sample : PB129155ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129155ZHE#08

Quant Time: May 30 06:57:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 29 02:14:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	260242	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	482043	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	487523	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	217120	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	171262	55.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.46%	
35) Dibromofluoromethane	7.55	113	126337	51.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.32%	
50) Toluene-d8	10.06	98	603590	53.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.52%	
62) 4-Bromofluorobenzene	12.38	95	226004	54.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.34%	
Target Compounds						
16) Acetone	3.78	43	10183	5.861	ug/l	93
43) Isopropyl Acetate	8.13	43	65834	10.647	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN052920\
Data File : VN061576.D
Acq On : 29 May 2020 18:43
Operator : JC/MD
Sample : PB129155ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129155ZHE#08

Quant Time: May 30 06:57:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052820W.M
Quant Title : SW846 8260
QLast Update : Fri May 29 02:14:10 2020
Response via : Initial Calibration

