

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052924\
 Data File : VN082373.D
 Acq On : 29 May 2024 16:16
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: May 30 05:45:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	189712	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	335377	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	313348	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	138166	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	153967	51.398	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.800%
35) Dibromofluoromethane	8.165	113	107844	52.273	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.540%
50) Toluene-d8	10.570	98	427080	53.994	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.980%
62) 4-Bromofluorobenzene	12.847	95	149149	52.830	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	105.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052924\
Data File : VN082373.D
Acq On : 29 May 2024 16:16
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: May 30 05:45:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

