

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052924\
 Data File : VN082378.D
 Acq On : 29 May 2024 18:17
 Operator : JC\MD
 Sample : P2612-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-203

Quant Time: May 30 05:47:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	180246	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	325982	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	316277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	135317	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	146685	51.538	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.080%	
35) Dibromofluoromethane	8.165	113	107139	53.428	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.860%	
50) Toluene-d8	10.571	98	419389	54.550	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 109.100%	
62) 4-Bromofluorobenzene	12.847	95	147054	53.590	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 107.180%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	82506	70.667	ug/l	100
20) Methylene Chloride	5.277	84	13324	5.656	ug/l #	77
25) 2-Butanone	7.494	43	12650	7.146	ug/l #	75
40) Benzene	8.612	78	16279	1.623	ug/l	99
43) Isopropyl Acetate	8.688	43	265700	39.840	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052924\
Data File : VN082378.D
Acq On : 29 May 2024 18:17
Operator : JC\MD
Sample : P2612-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-203

Quant Time: May 30 05:47:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

