

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052924\
 Data File : VN082381.D
 Acq On : 29 May 2024 19:30
 Operator : JC\MD
 Sample : P2612-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11933

Quant Time: May 30 05:48:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 16 05:21:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	181622	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	316943	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	314263	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	135109	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	146652	51.136	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.280%		
35) Dibromofluoromethane	8.171	113	104275	53.483	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.960%		
50) Toluene-d8	10.571	98	417299	55.826	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 111.660%		
62) 4-Bromofluorobenzene	12.853	95	147717	55.366	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 110.740%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	83592	71.055	ug/l	89
20) Methylene Chloride	5.289	84	11377	4.793	ug/l #	82
25) 2-Butanone	7.494	43	12860	7.210	ug/l #	87
40) Benzene	8.612	78	20349	2.087	ug/l #	89
43) Isopropyl Acetate	8.694	43	258910	39.929	ug/l #	74
51) 4-Methyl-2-Pentanone	10.447	43	19218	5.216	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN052924\
Data File : VN082381.D
Acq On : 29 May 2024 19:30
Operator : JC\MD
Sample : P2612-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-11933

Quant Time: May 30 05:48:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051524W.M
Quant Title : SW846 8260
QLast Update : Thu May 16 05:21:06 2024
Response via : Initial Calibration

