

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN053019\
 Data File : VN055893.D
 Acq On : 30 May 2019 12:08
 Operator : JC/SP
 Sample : K2642-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-02-052919

Quant Time: May 30 19:00:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	359889	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	551360	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	511446	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	173947	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	187577	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
35) Dibromofluoromethane	7.59	113	163471	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
50) Toluene-d8	10.09	98	660163	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
62) 4-Bromofluorobenzene	12.40	95	216017	47.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.22%	
Target Compounds						
16) Acetone	3.82	43	139451	85.356	ug/l	99
20) Methylene Chloride	4.54	84	5767	1.630	ug/l	95
43) Isopropyl Acetate	8.17	43	9891	1.268	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN053019\
Data File : VN055893.D
Acq On : 30 May 2019 12:08
Operator : JC/SP
Sample : K2642-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-02-052919

Quant Time: May 30 19:00:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

