

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN053019\
 Data File : VN055903.D
 Acq On : 30 May 2019 16:39
 Operator : JC/SP
 Sample : K3112-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPDES-1

Quant Time: May 31 08:30:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N052219W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 23 05:54:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	52228	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	284682	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	227536	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	107223	31.41	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	104.70%
60) 4-Bromofluorobenzene	12.40	95	81403	23.00	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	76.67%
63) Toluene-d8	10.09	98	326167	31.37	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	104.57%

Target Compounds

					Ovalue	
25) Chloroform	7.37	83	5704	1.061	ug/l	98
61) Tetrachloroethene	10.63	164	3234	1.050	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN053019\
Data File : VN055903.D
Acq On : 30 May 2019 16:39
Operator : JC/SP
Sample : K3112-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPDES-1

Quant Time: May 31 08:30:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N052219W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 23 05:54:54 2019
Response via : Initial Calibration

