

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
 Data File : VN077960.D
 Acq On : 30 May 2023 11:16
 Operator : JC\MD
 Sample : VN0530WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0530WBL01

Quant Time: May 31 02:03:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	405003	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	735954	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	638747	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	195791	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	282240	46.575	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.160%
35) Dibromofluoromethane	8.177	113	238689	49.736	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.480%
50) Toluene-d8	10.576	98	905405	49.015	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.040%
62) 4-Bromofluorobenzene	12.859	95	290552	44.536	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
Data File : VN077960.D
Acq On : 30 May 2023 11:16
Operator : JC\MD
Sample : VN0530WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0530WBL01

Quant Time: May 31 02:03:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

