

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
 Data File : VN077963.D
 Acq On : 30 May 2023 12:39
 Operator : JC\MD
 Sample : 02898-02DL 400X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW36BR-20230522DL

Quant Time: May 31 02:05:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	381417	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	672343	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	590334	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	182085	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	257073	45.045	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.100%
35) Dibromofluoromethane	8.177	113	221964	50.627	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.260%
50) Toluene-d8	10.571	98	837041	49.601	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.200%
62) 4-Bromofluorobenzene	12.859	95	269357	45.194	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.380%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.524	62	60139	10.989	ug/l	96
27) cis-1,2-Dichloroethene	7.500	96	353068	65.344	ug/l	98
44) Trichloroethene	9.359	130	139642	29.143	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
Data File : VN077963.D
Acq On : 30 May 2023 12:39
Operator : JC\MD
Sample : 02898-02DL 400X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW36BR-20230522DL

Quant Time: May 31 02:05:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

