

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
 Data File : VN077964.D
 Acq On : 30 May 2023 13:03
 Operator : JC\MD
 Sample : 02898-03DL 400X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW24BR-20230522DL

Quant Time: May 31 02:06:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	385843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	710631	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	638841	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	204428	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	269001	46.595	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.180%
35) Dibromofluoromethane	8.177	113	232427	50.157	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.320%
50) Toluene-d8	10.577	98	889246	49.856	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.720%
62) 4-Bromofluorobenzene	12.859	95	289274	45.921	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.840%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.524	62	8665	1.565	ug/l	93
27) cis-1,2-Dichloroethene	7.494	96	58005	10.612	ug/l	99
44) Trichloroethene	9.359	130	216844	42.817	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
Data File : VN077964.D
Acq On : 30 May 2023 13:03
Operator : JC\MD
Sample : 02898-03DL 400X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW24BR-20230522DL

Quant Time: May 31 02:06:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

