

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
 Data File : VN077966.D
 Acq On : 30 May 2023 13:51
 Operator : JC\MD
 Sample : 02973-11 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 INTERCEPTOR-SUMP-20230526

Quant Time: May 31 02:07:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	395196	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	711833	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	636649	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	196735	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	268861	45.468	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.940%
35) Dibromofluoromethane	8.177	113	229070	49.350	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.700%
50) Toluene-d8	10.577	98	905670	50.691	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.380%
62) 4-Bromofluorobenzene	12.853	95	295489	46.828	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.660%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.136	85	2965	0.659	ug/l	94
4) Vinyl Chloride	2.518	62	3102	0.547	ug/l #	68
12) 1,1-Dichloroethene	4.348	96	8828	2.124	ug/l #	83
27) cis-1,2-Dichloroethene	7.500	96	302454	54.025	ug/l	98
44) Trichloroethene	9.359	130	1792652	353.367	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
Data File : VN077966.D
Acq On : 30 May 2023 13:51
Operator : JC\MD
Sample : 02973-11 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
INTERCEPTOR-SUMP-20230526

Quant Time: May 31 02:07:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

