

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
 Data File : VN077967.D
 Acq On : 30 May 2023 14:15
 Operator : JC\MD
 Sample : 02973-13 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP-01-20230526

Quant Time: May 31 02:07:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	379091	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	702320	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	622577	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	194465	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	269024	47.429	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.860%
35) Dibromofluoromethane	8.177	113	223040	48.701	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.400%
50) Toluene-d8	10.577	98	860468	48.813	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.620%
62) 4-Bromofluorobenzene	12.859	95	281190	45.165	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.340%
Target Compounds						
44) Trichloroethene	9.353	130	6554	1.309	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
Data File : VN077967.D
Acq On : 30 May 2023 14:15
Operator : JC\MD
Sample : 02973-13 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP-01-20230526

Quant Time: May 31 02:07:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

