

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
 Data File : VN077975.D
 Acq On : 30 May 2023 17:27
 Operator : JC\MD
 Sample : 02911-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT149S1-20230522

Quant Time: May 31 02:10:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	375965	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	686100	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.876	117	603969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	190648	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	263592	46.857	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.720%		
35) Dibromofluoromethane	8.171	113	229555	51.309	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	102.620%		
50) Toluene-d8	10.577	98	854066	49.595	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.200%		
62) 4-Bromofluorobenzene	12.859	95	276317	45.432	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.860%		
Target Compounds						Qvalue
19) Methyl tert-butyl Ether	5.812	73	176084	12.302	ug/l	99
44) Trichloroethene	9.359	130	21450	4.387	ug/l	81
64) Tetrachloroethene	11.112	164	3980	0.942	ug/l	84
91) 1,2-Dichlorobenzene	14.112	146	1774	0.274	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
Data File : VN077975.D
Acq On : 30 May 2023 17:27
Operator : JC\MD
Sample : 02911-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT149S1-20230522

Quant Time: May 31 02:10:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

