

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
 Data File : VN077958.D
 Acq On : 30 May 2023 10:14
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 31 02:02:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.569	0.470	17.4	74	0.00
3 P	Chloromethane	0.682	0.526	22.9	81	0.00
4 C	Vinyl Chloride	0.717	0.672	6.3#	87	0.01
5 T	Bromomethane	0.531	0.506	4.7	94	0.01
6 T	Chloroethane	0.517	0.487	5.8	92	0.00
7 T	Trichlorofluoromethane	1.000	0.965	3.5	91	0.00
8 T	Diethyl Ether	0.365	0.347	4.9	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.553	0.534	3.4	92	0.00
10 T	Methyl Iodide	0.660	0.739	-12.0	105	0.00
11 T	Tert butyl alcohol	0.103	0.089	13.6	83	0.00
12 CM	1,1-Dichloroethene	0.526	0.527	-0.2#	94	0.00
13 T	Acrolein	0.073	0.049	32.9#	68	0.00
14 T	Allyl chloride	0.773	0.799	-3.4	105	0.01
15 T	Acrylonitrile	0.276	0.251	9.1	88	0.01
16 T	Acetone	0.235	0.232	1.3	100	0.00
17 T	Carbon Disulfide	1.417	1.314	7.3	92	0.00
18 T	Methyl Acetate	0.987	0.910	7.8	89	0.00
19 T	Methyl tert-butyl Ether	1.904	1.863	2.2	92	0.00
20 T	Methylene Chloride	0.705	0.629	10.8	97	0.00
21 T	trans-1,2-Dichloroethene	0.591	0.581	1.7	94	0.00
22 T	Diisopropyl ether	1.811	1.800	0.6	95	0.00
23 T	Vinyl Acetate	1.122	1.065	5.1	90	0.00
24 P	1,1-Dichloroethane	1.109	1.128	-1.7	96	0.01
25 T	2-Butanone	0.368	0.330	10.3	88	0.00
26 T	2,2-Dichloropropane	0.943	0.983	-4.2	93	0.00
27 T	cis-1,2-Dichloroethene	0.708	0.720	-1.7	96	0.00
28 T	Bromochloromethane	0.480	0.472	1.7	93	0.00
29 T	Tetrahydrofuran	0.241	0.210	12.9	86	0.00
30 C	Chloroform	1.220	1.212	0.7#	97	0.00
31 T	Cyclohexane	1.017	0.917	9.8	90	0.00
32 T	1,1,1-Trichloroethane	1.038	1.080	-4.0	96	0.00
33 S	1,2-Dichloroethane-d4	0.748	0.722	3.5	91	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.326	0.348	-6.7	94	0.00
36 T	1,1-Dichloropropene	0.477	0.502	-5.2	93	0.00
37 T	Ethyl Acetate	0.425	0.382	10.1	85	0.00
38 T	Carbon Tetrachloride	0.482	0.547	-13.5	99	0.00
39 T	Methylcyclohexane	0.584	0.564	3.4	85	0.00
40 TM	Benzene	1.407	1.478	-5.0	95	0.00
41 T	Methacrylonitrile	0.226	0.205	9.3	85	0.00
42 TM	1,2-Dichloroethane	0.511	0.516	-1.0	94	0.00
43 T	Isopropyl Acetate	0.694	0.688	0.9	91	0.00
44 TM	Trichloroethene	0.356	0.387	-8.7	99	0.00
45 C	1,2-Dichloropropane	0.359	0.385	-7.2#	96	0.00
46 T	Dibromomethane	0.257	0.266	-3.5	96	0.00
47 T	Bromodichloromethane	0.492	0.548	-11.4	98	0.00
48 T	Methyl methacrylate	0.329	0.312	5.2	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
 Data File : VN077958.D
 Acq On : 30 May 2023 10:14
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: May 31 02:02:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 16 04:07:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	89	0.00
50 S	Toluene-d8	1.255	1.283	-2.2	90	0.00
51 T	4-Methyl-2-Pentanone	0.413	0.397	3.9	87	0.00
52 CM	Toluene	0.876	0.964	-10.0#	97	0.00
53 T	t-1,3-Dichloropropene	0.521	0.579	-11.1	97	0.00
54 T	cis-1,3-Dichloropropene	0.574	0.633	-10.3	96	0.00
55 T	1,1,2-Trichloroethane	0.348	0.363	-4.3	95	0.00
56 T	Ethyl methacrylate	0.526	0.546	-3.8	91	0.00
57 T	1,3-Dichloropropane	0.596	0.615	-3.2	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.219	15.4	76	0.00
59 T	2-Hexanone	0.305	0.277	9.2	82	0.00
60 T	Dibromochloromethane	0.359	0.409	-13.9	99	0.00
61 T	1,2-Dibromoethane	0.347	0.369	-6.3	92	0.00
62 S	4-Bromofluorobenzene	0.443	0.450	-1.6	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.350	0.412	-17.7	107	0.00
65 PM	Chlorobenzene	1.084	1.173	-8.2	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.375	0.445	-18.7	101	0.00
67 C	Ethyl Benzene	1.942	2.132	-9.8#	95	0.00
68 T	m/p-Xylenes	0.736	0.816	-10.9	97	0.00
69 T	o-Xylene	0.724	0.793	-9.5	95	0.00
70 T	Styrene	1.150	1.283	-11.6	95	0.00
71 P	Bromoform	0.250	0.303	-21.2	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	4.310	4.660	-8.1	95	0.00
74 T	N-amyl acetate	1.593	1.426	10.5	82	0.00
75 P	1,1,2,2-Tetrachloroethane	1.326	1.281	3.4	90	0.00
76 T	1,2,3-Trichloropropane	1.252	1.143	8.7	84	0.00
77 T	Bromobenzene	0.984	1.017	-3.4	96	0.00
78 T	n-propylbenzene	5.081	5.333	-5.0	94	0.00
79 T	2-Chlorotoluene	3.032	3.199	-5.5	95	0.00
80 T	1,3,5-Trimethylbenzene	3.622	3.895	-7.5	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.392	0.472	-20.4	107	0.00
82 T	4-Chlorotoluene	3.017	3.056	-1.3	94	0.00
83 T	tert-Butylbenzene	3.162	3.307	-4.6	93	0.00
84 T	1,2,4-Trimethylbenzene	3.563	3.848	-8.0	96	0.00
85 T	sec-Butylbenzene	4.494	4.564	-1.6	89	0.00
86 T	p-Isopropyltoluene	3.583	3.805	-6.2	92	0.00
87 T	1,3-Dichlorobenzene	1.819	1.832	-0.7	93	0.00
88 T	1,4-Dichlorobenzene	1.805	1.783	1.2	91	0.00
89 T	n-Butylbenzene	3.296	3.091	6.2	83	0.00
90 T	Hexachloroethane	0.589	0.698	-18.5	99	0.00
91 T	1,2-Dichlorobenzene	1.699	1.799	-5.9	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.215	16.3	77	0.00
93 T	1,2,4-Trichlorobenzene	0.993	0.764	23.1	71	0.00
94 T	Hexachlorobutadiene	0.412	0.335	18.7	71	0.00
95 T	Naphthalene	3.418	2.249	34.2#	61	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN053023\
Data File : VN077958.D
Acq On : 30 May 2023 10:14
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: May 31 02:02:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N051523W.M
Quant Title : SW846 8260
QLast Update : Tue May 16 04:07:42 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.972	0.780	19.8	72	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6