



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN053023\  
Data File : VN077957.D  
Acq On : 30 May 2023 09:40  
Operator : JC\MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N051523W.M  
Title : SW846 8260  
Last Update : Tue May 16 04:07:42 2023



AutoFind: Scans 1853, 1854, 1855; Background Corrected with Scan 1844

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 19.8         | 18776      | PASS                |
| 75             | 95              | 30              | 60              | 53.8         | 51155      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 95037      | PASS                |
| 96             | 95              | 5               | 9               | 6.5          | 6187       | PASS                |
| 173            | 174             | 0.00            | 2               | 1.2          | 886        | PASS                |
| 174            | 95              | 50              | 100             | 76.6         | 72779      | PASS                |
| 175            | 174             | 5               | 9               | 7.0          | 5082       | PASS                |
| 176            | 174             | 95              | 101             | 98.3         | 71576      | PASS                |
| 177            | 176             | 5               | 9               | 5.9          | 4251       | PASS                |

Average of 12.853 to 12.865 min.: VN077957.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z   | abund. | m/z   | abund. | m/z   | abund. |
|-------|--------|-------|--------|-------|--------|-------|--------|
| 35.95 | 882    | 50.05 | 18776  | 63.00 | 4185   | 76.05 | 426    |
| 37.05 | 4790   | 51.00 | 5663   | 64.05 | 480    | 76.80 | 74     |
| 38.00 | 4383   | 51.80 | 74     | 65.05 | 569    | 77.70 | 10     |
| 39.00 | 1649   | 52.05 | 198    | 67.00 | 291    | 78.05 | 20     |
| 39.90 | 325    | 54.95 | 462    | 68.00 | 9986   | 78.90 | 4582   |
| 40.80 | 63     | 55.95 | 2019   | 69.00 | 10863  | 79.85 | 1079   |
| 43.95 | 623    | 57.00 | 3047   | 70.05 | 891    | 80.90 | 4035   |
| 44.90 | 1103   | 57.80 | 51     | 71.95 | 738    | 81.80 | 901    |
| 47.00 | 1007   | 60.00 | 1342   | 73.00 | 5599   | 86.10 | 132    |
| 47.90 | 656    | 61.00 | 5074   | 74.00 | 18697  | 86.95 | 3803   |
| 49.00 | 4093   | 62.00 | 5654   | 75.00 | 51155  | 88.00 | 3545   |

Average of 12.853 to 12.865 min.: VN077957.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z    | abund. | m/z    | abund. | m/z    | abund. |
|--------|--------|--------|--------|--------|--------|--------|--------|
| 90.85  | 559    | 106.70 | 58     | 129.90 | 537    | 147.90 | 168    |
| 92.00  | 3573   | 109.80 | 72     | 130.80 | 103    | 148.90 | 70     |
| 92.95  | 4766   | 110.80 | 119    | 134.90 | 195    | 149.60 | 145    |
| 94.00  | 13156  | 113.10 | 60     | 136.85 | 365    | 152.90 | 62     |
| 95.00  | 95037  | 114.85 | 157    | 140.95 | 1311   | 154.75 | 192    |
| 96.00  | 6187   | 115.90 | 385    | 141.80 | 61     | 156.80 | 60     |
| 97.00  | 248    | 116.85 | 875    | 142.80 | 707    | 161.10 | 52     |
| 103.80 | 664    | 117.90 | 256    | 143.00 | 396    | 171.75 | 280    |
| 104.80 | 105    | 118.75 | 563    | 145.00 | 81     | 172.15 | 388    |
| 105.75 | 501    | 127.85 | 253    | 146.00 | 121    | 173.05 | 886    |
| 106.00 | 143    | 128.80 | 71     | 147.10 | 327    | 174.00 | 72779  |

Average of 12.853 to 12.865 min.: VN077957.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z | abund. | m/z | abund. | m/z | abund. |
|--------|--------|-----|--------|-----|--------|-----|--------|
| 174.95 | 5082   |     |        |     |        |     |        |
| 175.95 | 71576  |     |        |     |        |     |        |
| 176.95 | 4251   |     |        |     |        |     |        |
| 178.00 | 52     |     |        |     |        |     |        |
| 207.00 | 78     |     |        |     |        |     |        |
| 209.20 | 61     |     |        |     |        |     |        |
| 281.20 | 245    |     |        |     |        |     |        |
| 327.10 | 55     |     |        |     |        |     |        |

Instrument :

MSVOA\_N

ClientSampleId :

BFB