

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN053119\
 Data File : VN055914.D
 Acq On : 31 May 2019 13:02
 Operator : JC/SP
 Sample : K3104-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RBR-251-688

Quant Time: Jun 01 02:14:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	386848	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	585783	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	532596	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	188453	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	195821	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
35) Dibromofluoromethane	7.59	113	170758	46.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.22%	
50) Toluene-d8	10.09	98	702072	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
62) 4-Bromofluorobenzene	12.40	95	225040	46.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.40%	
Target Compounds						
16) Acetone	3.82	43	17198	9.793	ug/l	99
20) Methylene Chloride	4.54	84	116441	30.616	ug/l	96
43) Isopropyl Acetate	8.17	43	10332	1.246	ug/l #	86
95) Naphthalene	15.13	128	13218	1.516	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN053119\
Data File : VN055914.D
Acq On : 31 May 2019 13:02
Operator : JC/SP
Sample : K3104-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RBR-251-688

Quant Time: Jun 01 02:14:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

