

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN053119\
 Data File : VN055920.D
 Acq On : 31 May 2019 15:33
 Operator : JC/SP
 Sample : K3134-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 01 02:33:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
 Quant Title : SW846 8260
 QLast Update : Fri May 24 06:41:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	352211	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	536745	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	483415	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	152710	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	181185	49.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.62%	
35) Dibromofluoromethane	7.59	113	156225	46.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.08%	
50) Toluene-d8	10.09	98	636039	48.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.72%	
62) 4-Bromofluorobenzene	12.40	95	199117	44.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.22%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN053119\
Data File : VN055920.D
Acq On : 31 May 2019 15:33
Operator : JC/SP
Sample : K3134-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 01 02:33:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N052319W.M
Quant Title : SW846 8260
QLast Update : Fri May 24 06:41:54 2019
Response via : Initial Calibration

